



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 25, 2011 INITIALS: JEC OWNER: WASHINGTON STATE PATROL
DISTRICT: 4 ADDRESS: I-90E MP300
LOCATION Spokane 64 -1 COUNTY: Spokane
MAKE: Toledo TYPE Trk Scale ACCURACY CLASS: IIII
MODEL: 8530 ID. NO: 4342378-4 CAPACITY: 100000 lbs. SECTIONS: 2
MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Sec 1	20,000	20,000	0				
Center	20,000	20,000	0				
Sec 2	20,000	20,000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: May 31, 2011
Print name: David Cromer

State of Washington
County of Shurston
Signed and sworn to (or affirmed before
me on May 31, 2011 by
Charice L Pidcock
(Signature of notary public)
Charice L Pidcock
Printed Name of Notary
My appointment expires 12/4/14

