



**WASHINGTON STATE PATROL**  
**SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: Aug 17, 2011 INITIALS: Dec OWNER: WASHINGTON STATE PATROL

DISTRICT: 7 ADDRESS: I-5 MP235

LOCATION: BowHill 33 E1 COUNTY: Skagit

MAKE: Toledo TYPE Trk Scale ACCURACY CLASS: III L

MODEL: 8530 ID. NO: 4409217-4SX CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

East 1

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20020	+20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Oct 6, 2011

Print name: David Cromer

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before me on October 6, 2011 by Charice L Pidcock  
 (Signature of notary public)  
Charice L Pidcock  
 Printed Name of Notary  
 My appointment expires 12/4/14

