



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Oct 13, 2011 INITIALS: Jec OWNER: WASHINGTON STATE PATROL

DISTRICT: 2 ADDRESS: I-5N MP141

LOCATION SeaTac NB 26 E1 COUNTY: King

MAKE: Toledo TYPE Trk Scale ACCURACY CLASS: IIII

MODEL: 8530 ID. NO: 4433540-4NY CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

East1

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	19980	-20				
CENTER	20000	20000	0				
SEC 2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Oct 17, 2011

Print name: David Cromer

State of Washington
County of Murston
Signed and sworn to (or affirmed before
me on October 17, 2011 by
Charice L. Pidcock
(Signature of notary public)
Charice L. Pidcock
Printed Name of Notary
My appointment expires 12/4/14

