



**WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: June 20, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 4 ADDRESS: I-90E MP231

LOCATION Tokio 68 EB -2 COUNTY: Adams

MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII

MODEL: IQ310 ID. NO: 91441406 CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10x10

2

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Sec 1	20,000	20,000	0				
Center	20,000	20,000	0				
Sec 2	20,000	19980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: June 21, 2010  
Print name: David Cromer

State of Washington  
County of Thurston  
Signed and sworn to (or affirmed before  
me on June 21, 2010 by

Brenda A. Biscay  
(Signature of notary public)  
Brenda A. Biscay  
Printed Name of Notary  
My appointment expires 3.12.12

