



WASHINGTON STATE PATROL

SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 14, 2011 INITIALS: Dec OWNER: WASHINGTON STATE PATROL
 DISTRICT: 4 ADDRESS: I-90E MP231
 LOCATION Tokio 68 EB -2 COUNTY: Adams
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII
 MODEL: IQ310 ID. NO: 91441406 CAPACITY: 100000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10x10

2

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Sec 1	20,000	20,000	0				
Center	20,000	20,000	0				
Sec 2	20,000	19980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: June 30, 2011
 Print name: David Cromer

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before me on June 30, 2011 by
Charice L. Pidcock
 (Signature of notary public)
Charice L. Pidcock
 Printed Name of Notary
 My appointment expires 12/4/14

