



**WASHINGTON STATE PATROL**  
**SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: June 22, 2011 INITIALS: [Signature] OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 6 ADDRESS: I-90E Bullfrog  
 LOCATION CleElum 52 EB COUNTY: Kititas  
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII  
 MODEL: IQ-355 ID. NO: 91441423 CAPACITY: 100000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	20,000	20,020	+20				
SEC 1	20,000	20,000	0				
SEC 2	20,000	20,000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: [Signature: David Cromer] Date: July 19, 2011  
 Print name: David Cromer

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before me on July 19, 2011 by  
Charice L. Pidcock  
 (Signature of notary public)  
Charice L. Pidcock  
 Printed Name of Notary  
 My appointment expires 12/4/14

