



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 27 2012 INITIALS: Dec OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 6 ADDRESS: I-90E Bullfrog  
 LOCATION: CleElum 52 EB COUNTY: Kittitas  
 MAKE: RiceLake TYPE: Trk Scale ACCURACY CLASS: III L  
 MODEL: IQ-355 ID. NO: 91441423 CAPACITY: 100000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	20,000	20,000	0				
SEC 1	20,000	20,000	0				
SEC 2	20,000	20,000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-1 and 200446-0-L3184-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: July 17, 2012  
 Print name: David Cromer

State of Washington

County of Thurston

Signed and sworn to (or affirmed before

me on July 17, 2012

Patricia A. Geppert  
 (Signature of notary public)

Patricia A. Geppert  
 Printed Name of Notary

My appointment expires Sept. 4, 2015

