



**WASHINGTON STATE PATROL**  
**SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: May 9, 2011 INITIALS: DEC OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 3 ADDRESS: SR12 MP308  
 LOCATION Wallula 43 COUNTY: WallaWalla  
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII  
 MODEL: IQ310 ID. NO: 92201007 CAPACITY: 100000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: ELECTRONIC PLATFORM 10x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	19980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: May 31, 2011  
 Print name: David Cromer

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before me on May 31, 2011 by  
Charice L. Pidcock  
 (Signature of notary public)  
Charice L. Pidcock  
 Printed Name of Notary  
 My appointment expires 12/4/14

