



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 24, 2011 INITIALS: JCS OWNER: WASHINGTON STATE PATROL
 DISTRICT: 4 ADDRESS: SR395 MP182
 LOCATION DeerPark 65 COUNTY: Spokane
 MAKE: Rice Lake TYPE Trk Scale ACCURACY CLASS: IIII
 MODEL: IQ310 ID. NO: 93495056 CAPACITY: 100000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 8x10

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|-------------|------------------|-------|---------------|-------------|------------------|-------|
| Sec 1 | 20,000 | 20,000 | 0 | | | | |
| Center | 20,000 | 20,000 | 0 | | | | |
| Sec 2 | 20,000 | 20,020 | +20 | | | | |
| | | | | | | | |
| | | | | | | | |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: May 31, 2011
 Print name: David Cromer

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before
 me on May 31, 2011 by
Charice L Pidcock
 (Signature of notary public)
Charice L Pidcock
 Printed Name of Notary
 My appointment expires 12/4/14

