



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: April 13, 2010 INITIALS: Dec OWNER: WASHINGTON STATE PATROL

DISTRICT: 1 ADDRESS: I-5 MP117

LOCATION FtLewis 7 E

COUNTY: Pierce

MAKE: RiceLake

TYPE Trk Scale

ACCURACY CLASS: IIII

MODEL: IQ 310

ID. NO: 96478582

CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs.

REG. ELEMENT: Digital

PLATFORM 11x18

EAST

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20020	+20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer

Date: June 21, 2010

Print name: David Cromer

State of Washington
County of Shurston
Signed and sworn to (or affirmed before
me on June 21, 2010 by

Brenda A. Biscay

(Signature of notary public)

Brenda A. Biscay

Printed Name of Notary

My appointment expires 3.12.12

