



**WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: June 19, 2012 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 1 ADDRESS: SR7 MP49

LOCATION: Elk Plain 16 COUNTY: Pierce

MAKE: RiceLake TYPE: Trk Scale ACCURACY CLASS: IIII

MODEL: IQ310 ID. NO: 96488779 CAPACITY: 100000 lbs. SECTIONS: 2


MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10'x12'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: June 14, 2012  
Print name: David Cromer



State of Washington  
County of Thurston  
Signed and sworn to (or affirmed before me on June 14, 2012 by

Patricia A. Geppert  
(Signature of notary public)  
Patricia A. Geppert

Printed Name of Notary  
My appointment expires Sept. 4, 2015