WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: April 19, 2011 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 1 ADDRESS: SR7 MP49

LOCATION Elk Plain 16 COUNTY: Pierce

MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIIL

MODEL: IQ310 ID. NO: 96488779 CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10'x12'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20020	+20				
CENTER	20000	20000	0				
SEC 2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: April 28, 2011

Print name: David Cromer

State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on April 28, 2011 by
Charice L Pidcock
(Signature of notary public)
Charice L. Pidcock
Printed Name of Notary
My appointment expires 12/4/14

CHARICE L. PIDCOCK
COMMISSION EXPIRES
NOTARY
PUBLIC
DEC. 4, 2014
STATE OF WASHINGTON

3000-362-021(R 09/06)