



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: May 13, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 4 ADDRESS: I-90W MP231
 LOCATION: Tokio 69 WB COUNTY: Adams
 MAKE: RiceLake TYPE: Trk Scale ACCURACY CLASS: III L
 MODEL: IQ310 ID. NO: 97120470 CAPACITY: 100000 lbs. SECTIONS: 1
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM: 8x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20,000	20,000	0				
CENTER	20,000	20,000	0				
SEC 2	20,000	20,020	+20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: June 21, 2010
 Print name: David Cromer

State of Washington
 County of Shurston
 Signed and sworn to (or affirmed before
 me on June 21, 2010 by

Brenda A. Biscay
 (Signature of notary public)
Brenda A. Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

