



WASHINGTON STATE PATROL

SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 14, 2011 INITIALS: Dec OWNER: WASHINGTON STATE PATROL
 DISTRICT: 4 ADDRESS: I-90W MP231
 LOCATION Tokio 69 WB COUNTY: Adams
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII
 MODEL: IQ310 ID. NO: 97120470 CAPACITY: 100000 lbs. SECTIONS: 1
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 8x10

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|-------------|------------------|-------|---------------|-------------|------------------|-------|
| SEC 1 | 20,000 | 20,000 | 0 | | | | |
| CENTER | 20,000 | 20,000 | 0 | | | | |
| SEC 2 | 20,000 | 20,000 | 0 | | | | |
| | | | | | | | |
| | | | | | | | |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: June 30, 2011
 Print name: David Cromer

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before me on June 30, 2011 by Charice L. Pidcock
 (Signature of notary public)
Charice L. Pidcock
 Printed Name of Notary
 My appointment expires 12/4/14

