



**WASHINGTON STATE PATROL**  
**SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: Nov 1, 2011 INITIALS: DEC OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 7 ADDRESS: SR20 MP68  
 LOCATION: Sedro W 30 COUNTY: Skagit  
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII  
 MODEL: IQ310 ID. NO: 97120551 CAPACITY: 100000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 8x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Nov 30, 2011

Print name: David Cromer

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before  
 me on Nov. 30-11 by  
Patricia A. Seppert  
 (Signature of notary public)  
Patricia A. Seppert  
 Printed Name of Notary  
 My appointment expires Sept 4, 2015

