



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: Aug 10, 2010 INITIALS: JEC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 8 ADDRESS: SR101 MP193
 LOCATION Forks 13 COUNTY: Clallum
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII
 MODEL: IQ310 ID. NO: 98025440 CAPACITY: 100000 lbs. SECTIONS: 3
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	19980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 24, 2010
 Print name: David Cromer

State of Washington
 County of Shenandoah
 Signed and sworn to (or affirmed before
 me on Aug 24, 2010 by

Brenda A Biscay
 (Signature of notary public)
Brenda A Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

