



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: May 3, 2011 INITIALS: Dec OWNER: WASHINGTON STATE PATROL
 DISTRICT: 3 ADDRESS: SR395N 10
 LOCATION N Pasco 49 COUNTY: Franklin
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: IIII
 MODEL: IQ310 ID. NO: 98025661 CAPACITY: 100000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: ELECTRONIC PLATFORM 10x14

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	19980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: May 31, 2011
 Print name: David Cromer

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before
 me on May 31, 2011 by
Charice L. Pidcock
 (Signature of notary public)
Charice L. Pidcock
 Printed Name of Notary
 My appointment expires 12/4/14

