



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: June 7, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 6 ADDRESS: SR2 MP105
 LOCATION Peshastin-Dryden 92 COUNTY: Chelan
 MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: III
 MODEL: IQ310 ID. NO: 98438244 CAPACITY: 100000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 8x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20,000	20,000	0				
CENTER	20,000	20,000	0				
SEC 2	20,000	19,980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: July 1, 2010
 Print name: David Cromer

State of Washington
 County of Shurston
 Signed and sworn to (or affirmed before
 me on July 1, 2010 by
Brenda A. Biscay
 (Signature of notary public)
Brenda A. Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

