



**WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: July 24, 2012 INITIALS: JCC OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 2 ADDRESS: I-5N MP141  
 LOCATION: SeaTac NB 26 W COUNTY: King  
 MAKE: Rice Lake TYPE: Trk Scale ACCURACY CLASS: IIII  
 MODEL: IQ310 ID. NO: 98438345 CAPACITY: 100000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10x12

West

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 20, 2012

Print name: David Cromer

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before  
 me on August 20, 2012 by

Patricia A. Geppert  
 (Signature of notary public)  
PATRICIA A. Geppert

Printed Name of Notary  
 My appointment expires Sept. 4, 2015

