



**WASHINGTON STATE PATROL**  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: April 27, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 3 ADDRESS: SR12 MP342  
 LOCATION WallaWalla 51 COUNTY: WallaWalla  
 MAKE: Fairbanks TYPE Trk Scale ACCURACY CLASS: III L  
 MODEL: CA ID. NO: G344903 CAPACITY: 50000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: Dial PLATFORM 10x12

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	4000	4000	0	CENTER	20000	20000	0
CENTER	8000	8000	0	SEC 1	20000	20000	0
CENTER	12000	12000	0	SEC 2	20000	20000	0
CENTER	16000	16000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: June 28, 2010  
 Print name: David Cromer

State of Washington  
 County of Shuster  
 Signed and sworn to (or affirmed before  
 me on June 28, 2010 by  
Brenda A. Biscay  
 (Signature of notary public)  
Brenda A. Biscay  
 Printed Name of Notary  
 My appointment expires 3-12-12

