WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 4, 2011 INITIALS: JEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 3 ADDRESS: SR12 MP342

LOCATION WallaWalla 51 COUNTY: WallaWalla

MAKE: Fairbanks TYPE Trk Scale ACCURACY CLASS: IIIL

MODEL: CA ID. NO: G344903 CAPACITY: 50000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: Dial PLATFORM 10x12

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	4000	4000	0	CENTER	20000	20000	0
CENTER	8000	8000	0	SEC 1	20000	20000	0
CENTER	12000	12000	0	SEC 2	20000	20000	0
CENTER	16000	16000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: May 31, 2011

Print name: David Cromer

State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on May 31, 2011 by
Charice L Pidcock
(Signature of notary public)
Charice L Pidcock
Printed Name of Notary
My appointment expires 12/4/14

CHARICE L. PIDCOCK
COMMISSION EXPIRES
NOTARY
PUBLIC
DEC. 4, 2014
STATE OF WASHINGTON

3000-362-021(R 09/06)