



WASHINGTON STATE PATROL

SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 27, 2012 INITIALS: DC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 3 ADDRESS: SR97 MP56
 LOCATION: Toppenish 42 COUNTY: Yakima
 MAKE: Fairbanks TYPE: Trk Scale ACCURACY CLASS: IIII
 MODEL: 42-9000-BW ID. NO: G608891 CAPACITY: 62500 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: Dial PLATFORM 8x12

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	4000	4000	0	CENTER	20000	20000	0
CENTER	8000	8000	0	SEC 1	20000	20000	0
CENTER	12000	12000	0	SEC 2	20000	20000	0
CENTER	16000	16000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-1 and 200446-0-L3184-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer

Date: July 17, 2012

Print name: David Cromer

State of Washington
 County of Thurston

Signed and sworn to (or affirmed before me on July 17, 2012)

Patricia A. Geppert
 (Signature of notary public)

PATRICIA A. GEPPERT
 Printed Name of Notary

My appointment expires Sept. 4, 2015

