



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: July 29, 2010 INITIALS: DC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 1 ADDRESS: SR410 MP19
 LOCATION Buckley 18 COUNTY: Pierce
 MAKE: Fairbanks TYPE Trk Scale ACCURACY CLASS: IIIIL
 MODEL: 42-9000 ID. NO: G673188 CAPACITY: 62500 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: Dial PLATFORM 8x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	4000	4000	0	CENTER	20000	20000	0
CENTER	8000	8000	0	SEC 1	20000	20000	0
CENTER	12000	12000	0	SEC 2	20000	20000	0
CENTER	16000	16000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 4, 2010
 Print name: David Cromer

State of Washington
 County of Shurston
 Signed and sworn to (or affirmed before
 me on Aug 4, 2010 by
Brenda A Biscay
 (Signature of notary public)
Brenda A. Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

