

# WASHINGTON STATE PATROL

## Report of Calibration

Guth-Eutechnics Digital Thermometer

Guth Part Number: GU-4300-DT

Serial Number: 091796

The Digital Thermometer calibration was verified at 30.00°C, 34.00°C, and 37.00°C by comparison to temperature standards traceable to the National Institute of Standards & Technology.

Simulator Temperature Degrees C	Master Digital Thermometer S/N 082708	Digital Thermometer Reading	Test Date
30.00	30.01°C	30.00°C	06/08/01
34.00	34.00°C	33.99°C	06/08/01
37.00	36.99°C	36.98°C	06/08/01

Tested by: TR [Signature]

Date: 6-8-01

Bostec Inc.