



# Calibration Certificate

<b>Description:</b>	MULTIMETER	<b>Certificate Number:</b>	649084-69920523:1207138153
<b>Manufacturer:</b>	FLUKE	<b>Date of Calibration:</b>	02 April 2008
<b>Model:</b>	70 III	<b>Date of Certificate:</b>	02 April 2008
<b>Serial Number:</b>	69920523	<b>Date Due:</b>	02 April 2009
<b>Customer Name:</b>	STATE OF WASHINGTON	<b>Procedure Name:</b>	MFG MANUAL
<b>City, State:</b>	SEATTLE, WA	<b>Procedure Revision:</b>	7/98
<b>Customer Item ID:</b>	69920523	<b>Data Type:</b>	FOUND-LEFT
<b>PO Number:</b>	WA ST PATROL CCS	<b>Temperature:</b>	23 ± 3.0 °Celsius
<b>RMA Number:</b>	3942977	<b>Relative Humidity:</b>	25% ≤ RH ≤ 60%
		<b>Test Result:</b>	PASS

The Fluke Corporation, NQA ISO 9001:2000 ISO Certification No. 10100/2, certifies that the instrument identified above was calibrated in accordance with applicable Fluke calibration procedures. Its calibration processes are ISO-9001 controlled and are designed to certify that the instrument was within its published specifications at the time of calibration.

The measurement standards and instruments used during the calibration of this instrument are traceable to the United States National Institute of Standards and Technology (NIST), other reputable National Institutes, natural physical constants, consensus standards, or by ratio type measurements.

This certificate applies to only the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. The user is obliged to have the in object recalibrated at appropriate intervals. Calibration Certificates without signature are not valid.

The Data type that could be found in this certificate is interpreted as follows:

- As Found — The unit needed adjustment and/or repair.
- As Left — The unit was adjusted and/or repaired.
- As Found/ As Left — The unit was calibrated without any adjustment and/or repair performed.

**Comments:**

Long Le

Metrology Technician

**Standards Used**

Asset #	Instrument Model	Cal Date	Cal Due
10127	FLUKE 5520A CALIBRATOR	28 November 2007	28 August 2008

End of Report