

Certificate of Calibration Everett Service Center



ISO 9001:2008 (10100/2)

Certificate Number: 649084-708600863:1374406014

Result Summary: PASS.	Date of Calibration: 21 July 2013
Data Type: FOUND-LEFT	Recommended Due Date: 21 July 2014
Manufacturer: FLUKE	Date of Certificate: 21 July 2013
Model: 70 III	Received Date: 19 July 2013
Serial Number: 708600863	Temperature: 23.5 °C
Description: MULTIMETER	Relative Humidity: 52 %RH
Procedure Name: FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision: 1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: 708600863
PO Number: WSP CCS
RMA Number: 30328215

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

A handwritten signature in cursive script, appearing to read 'Sierra Freeman'.

Sierra Freeman
Metrology Technician

FLUKE ®

Certificate Number:
649084-708600863:1374406014

Calibration Date:
21-Jul-13

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
10418	FLUKE 5520A CALIBRATOR	09 June 2013	09 October 2013

End of Report