

Certificate of Calibration Everett Service Center



ISO 9001:2008 (10100/2)

Certificate Number: 649084-708601073:1344442748

| | | | |
|-----------------|--|-----------------------|----------------|
| Result Summary: | PASS | Date of Calibration: | 08 August 2012 |
| Data Type: | FOUND-LEFT | Recommended Due Date: | 08 August 2013 |
| Manufacturer: | FLUKE | Date of Certificate: | 08 August 2012 |
| Model: | 70 III | Received Date: | 02 August 2012 |
| Serial Number: | 708601073 | Temperature: | 23.9 °C |
| Description: | MULTIMETER | Relative Humidity: | 43 %RH |
| Procedure Name: | FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A | Procedure Revision: | 1.0 |

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: 708601073
PO Number: CCS IMPAIRED DRIVING 5
RMA Number: 30086547

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 114 08/09 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

A handwritten signature in cursive script that reads 'Sierra Freeman'.

Sierra Freeman
Metrology Technician

FLUKE ®

Certificate Number:
649084-708601073:1344442748

Calibration Date:
8-Aug-12

Standards Used

| Asset # | Instrument Model | Cal Date | Cal Due |
|---------|------------------------|-------------|------------------|
| 13608 | FLUKE 5520A CALIBRATOR | 17 May 2012 | 17 February 2013 |

End of Report