



ISO 9001:2008 (10100/2)

Certificate of Calibration

Everett Service Center

Certificate Number: 649084-72830395:1368043676

Result Summary: PASS	Date of Calibration: 08 May 2013
Data Type: FOUND-LEFT	Recommended Due Date: 08 May 2014
Manufacturer: FLUKE	Date of Certificate: 08 May 2013
Model: 70 III	Received Date: 02 May 2013
Serial Number: 72830395	Temperature: 22.5 °C
Description: MULTIMETER	Relative Humidity: 37 %RH
Procedure Name: FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision: 1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: 72830395
PO Number: CCS WASHINGTON STATE PATROL
RMA Number: 30277461

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

Corey Dinardi
Metrology Technician

FLUKE®

Certificate Number:
649084-72830395:1368043676

Calibration Date:
8-May-13

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
13677	FLUKE 5520A CALIBRATOR	17 November 2012	17 November 2013

End of Report