



STATE OF WASHINGTON
WASHINGTON STATE PATROL

811 East Roanoke Street • Seattle WA 98102 • (206) 720-3018 • www.wsp.wa.gov

June 27, 2022

Washington Association of Prosecuting Attorneys
206 10th Ave SE
Olympia, WA 98501

Subject: UPDATE: Potential Impeachment Disclosure Concerning Evidential Breath Test Results

This is an update to the previous notification(s) provided on June 16, 2021 and July 8, 2021.

On June 23, 2022, the Impaired Driving Section (IDS) identified three (3) additional breath tests which appear to have met the acceptance criteria outlined in WAC 448-16-060 but were not accepted by the Draeger Alcotest 9510. No breath test tickets were produced in these three cases as a result.

To date, no accepted breath tests have been identified as incorrectly meeting the acceptance criteria.

Below are the additional (three) identified breath tests which appear to have met the acceptance criteria in the WAC but were not accepted by the instrument:

Serial Number	Date	Obs Time	Citation	Operator	BrAC 1 IR	BrAC 1 EC	B1 Time	BrAC 2 IR	BrAC 2 EC	B2 Time	Sta	Truncated MEAN, 4	Truncated Mean -10%	Truncated Mean +10%	Rounded MEAN, 4	Rounded Mean -10%	Rounded Mean +10%
ARAF-0003	10/15/2021	13:50	21S09376	FLORES, ROBERTO DANIEL D	0.013	0.014	14:25	0.011	0.013	14:29	10	0.0127	0.011	0.013	0.0128	0.011	0.014
ARHC-0003	3/12/2022	03:24	1-22-000215	WALTERS, CHEYANNE M	0.019	0.017	03:55	0.017	0.016	03:59	10	0.0172	0.015	0.018	0.0173	0.015	0.019
ARHC-0003	3/12/2022	03:24	1-22-000215	WALTERS, CHEYANNE M	0.014	0.012	04:10	0.013	0.012	04:14	10	0.0127	0.011	0.013	0.0128	0.011	0.014

If you need additional information, please contact Lieutenant Jeffrey Leonard at Jeffrey.Leonard@wsp.wa.gov

Sincerely,

Lieutenant Jeffrey N. Leonard
Impaired Driving Section

JNL:jnl

cc: Dr. Fiona J. Couper, Forensic Laboratory Services Bureau
Ms. Amanda Black, Toxicology Laboratory Division

