

JAY INSLEE
Governor



JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL

811 East Roanoke Street • Seattle WA 98102 • (206) 720-3018 • www.wsp.wa.gov

May 2, 2022

Washington Association of Prosecuting Attorneys
206 10th Ave SE
Olympia, WA 98501

Subject: Potential Impeachment Disclosure Concerning Preliminary Breath Test Certifications

On April 21, 2022, the Impaired Driving Section (IDS) discovered that we failed to recertify Trooper Aaron Trombley as qualified to certify Preliminary Breath Test instruments (PBTs) for use pursuant to Washington Administrative Code 448-15. Based on IDS records, Trooper Trombley's PBT Technician certification expired on August 18, 2020.

A comprehensive review of records revealed that 21 PBT's, assigned to 20 individual troopers are affected. The attached document lists each affected trooper along with a corresponding PBT serial number. The document provides disclosure for PBTs certified after September 1, 2020 to April 27, 2022. Only PBTs certified by Trooper Trombley are affected.

IDS staff is currently re-certifying all affected PBTs.

If you need additional information, please contact Lieutenant Clark Jones at Clark.Jones@wsp.wa.gov.

Sincerely,

Lieutenant Clark G. Jones
Impaired Driving Section

CGJ:jnl
Attachment

Cc: Ms. Amanda Black, Toxicology Laboratory Division
Captain Dennis Bosman, Commercial Vehicle Enforcement Bureau
Dr. Fiona J. Couper, Forensic Laboratory Services Bureau
Captain Shane Nelson, Yakima



Breath Test Program

PBT Certification Record

Trooper Badge Number: 719 Dedrick Joliff

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
144855		1/12/2022	1309737	0.080	4/8/2023	0.081	No	0.000	Brad Neff
719		5/5/2021	1171207	0.080	7/10/2022	0.081	No	0.000	Aaron Trombley

Trooper Badge Number: 585 Cory Litchfield

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
116681		12/27/2021	1309737	0.080	4/8/2023	0.075	Yes	0.079	Brad Neff
116681		5/5/2021	1171207	0.080	7/10/2022	0.074	Yes	0.079	Aaron Trombley
116681		11/23/2020	1040427	0.080	11/5/2021	0.079	No	0.000	Brad Neff

Trooper Badge Number: 1021 Michael Benavidez

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
1266667		10/27/2021	1309737	0.080	4/8/2023	0.079	No	0.000	Aaron Trombley
126667		5/5/2021	1171207	0.080	7/10/2022	0.076	Yes	0.079	Aaron Trombley
126667		5/5/2021	1171207	0.080	7/10/2022	0.076	Yes	0.078	Aaron Trombley

Trooper Badge Number: 1231 Ryan Gustin

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
145810		5/5/2021	1171207	0.080	7/10/2022	0.082	Yes	0.080	Aaron Trombley
1455810		11/18/2020	1040427	0.080	11/5/2021	0.076	Yes	0.081	Aaron Trombley

Trooper Badge Number: 313 David Brandt

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
112380		10/27/2021	1309737	0.080	4/8/2023	0.074	Yes	0.082	Aaron Trombley
113380		5/5/2021	1171207	0.080	7/10/2022	0.078	No	0.000	Aaron Trombley
112380		12/2/2020	1040427	0.080	11/5/2021	0.089	Yes	0.080	Brad Neff

Trooper Badge Number: 185 Brian Mihelich

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
1084942		5/5/2021	1171207	0.080	7/10/2022	0.072	Yes	0.080	Aaron Trombley
1084942		12/2/2020	1040427	0.080	11/5/2021	0.079	No	0.000	Brad Neff

Trooper Badge Number: 810 Robert Morris

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
005913		10/27/2021	1309737	0.080	4/8/2023	0.070	Yes	0.082	Aaron Trombley
005913		3/25/2021	1308463	0.077	4/6/2023	0.085	Yes	0.077	Justin Rumsey
005913	00006890	10/13/2020	1171207	0.080	7/10/2022	0.081	No	0.000	Justin Rumsey
005913	00006890	9/7/2020	1171207	0.080	7/10/2022	0.084	Yes	0.080	Justin Rumsey

Trooper Badge Number: 934 Martin Finan

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
027332		6/15/2021	1309737	0.080	4/8/2023	0.067	Yes	0.080	Brad Neff
027332		12/2/2020	1040427	0.080	11/5/2021	0.088	Yes	0.082	Aaron Trombley

Trooper Badge Number: 1042 Jarryd Bivins

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
057883		4/25/2022	wo325341	0.077	11/18/2022	0.077	No	0.000	Justin Rumsey
057883		12/7/2021	1309737	0.080	4/8/2023	0.089	Yes	0.077	Aaron Trombley
057883		1/15/2021	1040427	0.080	11/5/2021	0.085	Yes	0.080	Brad Neff
057883		9/21/2020	925605	0.080	3/19/2021	0.079	Yes	0.087	Aaron Trombley

Trooper Badge Number: 529 Matthew Parham

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
033842		4/25/2022	wo325341	0.077	11/18/2024	0.079	No	0.000	Justin Rumsey
033842		12/7/2021	1309737	0.080	4/8/2023	0.076	Yes	0.078	Aaron Trombley
033842		1/15/2021	1040427	0.080	11/5/2021	0.084	Yes	0.080	Brad Neff

Trooper Badge Number: 700 Carlos Mata

Start Date: 9/1/2020

SerialNumber	TagNumber	CertificationDate	Gas Standard Number	GasStandard Value	Gas StandardExpiration Date	Results	Recalibrated	Recalibrated Results	Technician
063894		4/25/2022	wo325341	0.077	11/18/2024	0.076	No	0.000	Justin Rumsey
063894		12/7/2021	1309737	0.080	4/8/2023	0.076	Yes	0.077	Aaron Trombley
063894		7/16/2021	1309737	0.080	4/8/2023	0.080	No	0.000	Brad Neff
063894		1/15/2021	1040427	0.080	11/5/2021	0.077	Yes	0.080	Brad Neff

Trooper Badge Number: 1067 Thomas Dylan

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
144812		4/25/2022	wo325341	0.077	11/18/2024	0.078	No	0.000	Justin Rumsey
144812		12/7/2021	1309737	0.080	4/8/2023	0.079	No	0.000	Aaron Trombley
144812		1/15/2021	1040427	0.080	11/5/2021	0.081	No	0.000	
144812		1/15/2021	1040427	0.080	11/5/2021	0.081	No	0.000	Brad Neff

Trooper Badge Number: 457 Joseph DePalma

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
018280		4/25/2022	wo325341	0.077	11/18/2024	0.081	No	0.000	Justin Rumsey
018280		12/7/2021	1309737	0.080	4/8/2023	0.081	No	0.000	Aaron Trombley
032464		6/15/2021	1309737	0.080	4/8/2023	0.078	No	0.000	Brad Neff
032464		1/21/2021	1040427	0.080	11/5/2021	0.083	Yes	0.080	Brad Neff

Trooper Badge Number: 975 Chadwick Williams

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
132529		12/9/2021	1309737	0.080	4/8/2023	0.077	No	0.000	Aaron Trombley
132529		7/23/2021	1309737	0.080	4/8/2023	0.080	No	0.000	Aaron Trombley
132529		1/15/2021	1040427	0.080	11/5/2021	0.080	No	0.000	Brad Neff

Trooper Badge Number: 166 Daniel Munder

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
092407		12/6/2021	1309737	0.080	4/8/2023	0.081	No	0.000	Aaron Trombley
092407		9/16/2021	1309737	0.080	4/8/2023	0.080	No	0.000	Justin Destito
092407		6/1/2021	1428765	0.077	12/31/2023	0.069	Yes	0.078	Justin Rumsey

Trooper Badge Number: 539 Daniel Mosqueda

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
144813		10/27/2021	1309737	0.080	4/8/2023	0.088	Yes	0.079	Aaron Trombley
053910		12/1/2020	1040427	0.080	11/5/2021	0.035	Yes	0.081	Aaron Trombley
05910		12/1/2020	1040427	0.080	11/5/2021	0.035	Yes	0.081	Aaron Trombley

Trooper Badge Number: 1068 Andres Martinez

Start Date: 9/1/2020

SerialNumber	TagNumber	CertificationDate	Gas Standard Number	GasStandard Value	Gas StandardExpiration Date	Results	Recalibrated	Recalibrated Results	Technician
022851		5/5/2021	1171207	0.080	7/10/2022	0.080	No	0.000	Aaron Trombley
033851		12/2/2020	1040427	0.080	11/5/2021	0.081	No	0.000	Brad Neff

Trooper Badge Number: 1164 Chris Thorson

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
008660		4/26/2022	wo325341	0.077	11/18/2024	0.083	No	0.000	Justin Rumsey
008660	101094	9/8/2021	1309737	0.080	4/8/2023	0.074	Yes	0.080	Brad Neff
008660		12/2/2020	1040427	0.080	11/5/2021	0.091	Yes	0.080	Aaron Trombley

Trooper Badge Number: 308 Aaron Trombley

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
064465		10/27/2021	1309737	0.080	4/8/2023	0.073	Yes	0.081	Aaron Trombley
064465		12/1/2020	1040427	0.080	11/5/2021	0.077	No	0.000	Aaron Trombley

Trooper Badge Number: 1178

Start Date: 9/1/2020

Serial Number	Tag Number	Certification Date	Gas Standard Number	Gas Standard Value	Gas Standard Expiration Date	Results	Recalibrated	Recalibrated Results	Technician
144820		9/21/2020	925605	0.080	3/19/2021	0.080	No	0.000	Aaron Trombley