

Kendall

To: Ms. Jennifer Rice
Washington State Toxicology Lab
2203 Airport Way S. Ste 360
Seattle, WA 98134

RECEIVED

APR 11 2011

WSP TOX LAB

Fax: 206-262-6145


CERTIFICATE OF COMPLIANCE

This is to certify that the products listed below are in compliance with the current FDA Quality System Requirements (QSR) as stipulated in 21 CFR Part 820. Representative product was inspected and tested in accordance with current Kendall specifications and quality requirements.

<u>Product #</u>	<u>Description</u>	<u>Lot #</u>
8881352788	BCS GRA 16X100 10ML P.O+F	028502


This tube was designed for laboratory procedures requiring plasma or whole blood and chemistry procedures where glycolytic inhibition of the specimen is required. The tube was manufactured to the following specifications:

	<u>Range</u>	<u>Nominal</u>
Potassium Oxalate	15.0-26.0mg	20mg
Sodium Fluoride	20.0-32.0mg	25mg
Draw volume	9.0-11.0ml	10ml



David A. Olson
Vice President Regulatory Affairs

On this 10th day of April 2011, before me the undersigned notary public, personally appeared David Olson, proved to me through satisfactory evidence of identification, which were driver's license to be the person whose name is signed on the proceeding or attached document in my presence.



Notary Public

My commission expires:

11-03-2017

Date: March 31, 2011